

SIEVING TRIALS RESULT SHEET

CUSTOMER											
DATE	4/4/2006	TEST No:		1			SHEET No:		1		
SIEVE Dia.	200 mm	MACHINE TYPE	Meinzer 2								
SETTINGS	AMPLITUDE				OPERATION		CONTINUOUS			INTERMITTENT	
SAMPLE TYPE	Caffè Starbucks Coffee Ground				SAMPLE WEIGHT gms.			89.1g			
MACHINE FEATURES											

Running Time mins.	5 Min.				
Sieve size mm.	Actual Wt gms.	Individual % Retained	Cumulative % Retained		Cumulative % Passing
1.70	0.01	0	0		0
1.18	1.37	2	2		98
850um	18.23	20	22		78
600um	37.26	42	64		36
425um	13.80	15	79		21
300um	14.81	17	96		4
Receiver	3.90	4	100		0
TOTAL	89.4	100			

SIEVING TRIALS RESULT SHEET

CUSTOMER											
DATE	4/4/2006	TEST No:		2			SHEET No:		1		
SIEVE Dia.	200 mm	MACHINE TYPE	ROTAP								
SETTINGS	AMPLITUDE				OPERATION		CONTINUOUS			INTERMITTENT	
SAMPLE TYPE	Caffè Starbucks Coffee Ground				SAMPLE WEIGHT gms.			89.1g			
MACHINE FEATURES											

Running Time mins.	5 Min.				
Sieve size mm.	Actual Wt gms.	Individual % Retained	Cumulative % Retained		Cumulative % Passing
1.70	0.05	0	0		0
1.18	0.84	1	1		99
850um	17.09	19	20		80
600um	41.01	46	66		34
425um	14.81	17	83		17
300um	8.84	10	93		7
Receiver	6.50	7	100		0
TOTAL	89.1	100			